Research on rapid semi-quantitative determination of cadmium in pyrotechnics used for fireworks and firecrackers

Junyi Wu

Beihai Customs Comprehensive Technology Service Center, Beihai, Guangxi, China johnny2000@vip.qq.com

Johnny2000@vip.qq.com

Keywords: semi-quantitative; cadmium; pyrotechnics; fireworks and firecrackers.

Abstract: This study discloses a method for rapid semi-quantitative determination of cadmium in pyrotechnic powder for fireworks and firecrackers based on energy dispersive X-ray fluorescence spectrometer (EDXRF), including the following steps: preparation of samples, establishment of testing methods, determination of the characteristic line fluorescence intensity values of Cd element in samples, and according to the measured fluorescence intensity values, the cadmium content in the pyrotechnic powder samples can be semi-quantitatively determined. The method of the study has the advantages that: (1) the method is simple to operate, and the method can be repeatedly called for testing. Only one new test method needs to be built before the sample test. After the method is established, the test can be repeated at different times without re-establishing the test method for each test. After the first establishment of the new test method, the entire test process only includes three steps: sample preparation, sample loading into the sample cup and on-board testing. (2) The testing time is extremely short. After the sample is prepared, the entire measurement process takes only about 2 minutes. (3) Labor intensity is very low and the requirements for operators are not high. (4) The method has good stability, good repeatability and high credibility.

1. Introduction

In the prior art, no method for quickly semi-quantitatively or quantitatively determining the cadmium content in pyrotechnics used for fireworks and firecrackers based on energy dispersive X-ray fluorescence spectroscopy has been found, and we can find some standard method for instrumental analysis of cadmium in electronic products in electronic product testing standards. The currently used method standard is "Rapid Screening X-ray Fluorescence Spectrometry for Lead, Mercury, Chromium, Cadmium and Bromine in Restricted Substances in Electrical and Electronic Products" (GB/Z 21277-2007), and the analytical instrument used in this standard method is a wavelength dispersive X-ray fluorescence spectrometer whose fundamental principles of detection are quite different from the principles described herein.

The methods currently developed by energy dispersive X-ray fluorescence spectroscopy are mostly used for non-destructive qualitative analysis of samples. For semi-quantitative and quantitative elemental detection of solid samples, most samples are directly determined by powder tableting or melting methods, such as: "Determining the content of precious metals by X-ray fluorescence spectrometry" (GB/T 18043-2008) using non-destructive Detection method, "Alumina chemical analysis method and physical property measurement method Part 30 X-ray fluorescence spectrometry determination of element content" (GB/T6609.30-2009) using the melting method, "EDXRF method for direct determination of W-Fe- Ni-Co Alloy Mixture Components ("Nuclear Electronics and Detection Technology" Issue 5,2007) uses the tabletting method, "X-ray fluorescence spectrometry for rapid determination the contents of potassium, sodium, calcium and magnesium in potassium chloride products". (Analytical Instruments, Issue 6, 2013) using the tabletting method, "Rapid screening X-ray fluorescence spectrometry for lead, mercury, chromium, cadmium and bromine in electronic and electrical products" (GB/Z 21277) -2007) uses the tableting method and explosive, it is impossible to tablet or melt the pyrotechnic powder for the sample preparation.

So far, there has been no published literature report on the method of rapid semi-quantitative determination of cadmium content in pyrotechnics for fireworks and firecrackers based on energy dispersive X-ray fluorescence spectrometry.

2. Theory

After the pyrotechnic sample is excited by X-rays, different elements in the sample emit different characteristic X-rays. These characteristic lines are fingerprint information that identifies the target element in the sample. By measuring the characteristic X-ray fluorescence intensity of the target element in the sample, it is possible to semi-quantitatively analyze the amount of cadmium in the unknown pyrotechnic sample. The method directly uses the pyrotechnic powder sample of the fireworks and firecrackers to establish a specific analysis method, and semi-quantitatively analyzes the cadmium content in the sample according to the X-ray fluorescence intensity value of the characteristic line of the cadmium element.

3. Experiment section

3.1 Instrument and apparatus

Oven with accuracy to $\pm 2^{\circ}$ C. Analytical balance with accuracy to 0.1 mg. energy dispersive X-ray fluorescence spectrometer (EDXRF): United States Thermo Fisher (former Thermo Electron Corporation) Company QUANT'X series.

3.2 Operation step

(1) 10 to 30 g of the 40-100 mesh sieve sample powder is thoroughly mixed, placed in an oven, dried, placed in a desiccator and cooled to room temperature, and ready to be used.

(2) Weigh the sample of about 2 g, make sure the thickness of the powder sample in the sample cup is \geq 3mm.

(3) Gently tamper the sample cup 3 times on the hard ground and put the cup in the testing tank.

(4) Set the parameters of the EDXRF instrument as shown in Table 1.

Filter	Thin Cu	
Collimator	8.8mm	
Voltage	38v	
Electric current	Auto	
Analysis time	30s	
Count rate	Medium	
Atmosphere	Air	
Matrix effects	Not considered	
Energy range	0~40kev	
Analysis technique	Intensity correction	
sample thickness	≥3mm	

Table 1 Parameters of the EDXRF instrument

(4) Sample determination: determine the fluorescence intensity of the target element of the sample under the best analysis condition and read the values of it.

(5) $\omega = \frac{F}{C}$ Calculate the amount of cadmium in the sample by calculating the fluorescence intensity value according to the following formula:

Wherein, ω indicats the mass percentage of cadmium in the sample, the unit is %; C is the fluorescence intensity value of the corresponding characteristic line when the mass percentage of cadmium in the sample is 1%, and the unit is cps/mA; F is the fluorescence intensity value of the characteristic line of cadmium element measured by the energy dispersive X-ray fluorescence spectrometer in units of cps/mA.

4. Results and Discussion

4.1 Sample size and particle size

In the method, 10 to 30 g of the 40-100 mesh sieve sample powder is thoroughly mixed, placed in an oven, dried, placed in a desiccator and cooled to room temperature, and ready to be used. The reason why the particle size of the sample is set to $10 \sim 30$ g is that in the actual production process, the quality of the pyrotechnic powder for fireworks and firecrackers is uneven and the density of the cadmium powder is high, if the sample size is too small, the sample would not be representative and would be difficult to meet the requirements of the sample thickness in the sample cup which is required over 3mm thickness, and it will directly affect the accuracy of the test results. If the sample size is too large, it will affect the efficiency of the sample preparation. There are two main reasons why the sample must be passed through a 40-100 mesh sieve: Firstly, The energy dispersive X-ray fluorescence spectrometer analyzes the surface of the sample to get the fluorescence intensity of the characteristic line of cadmium element, if the sample with uneven particle size is likely to have a large particle size effect which would seriously affect the accuracy of the test results. So it must be sure to make the particle size of the sieved sample not to be too big to avoid increasing unevenness of particle size of the sample. A large amount of experimental data indicates that the particle size of the sieved sample is less than 40 mesh would cause little particle size effects. Secondly, if the pyrotechnic powder sample passes through a sieve of more than 100 mesh, the particle size will become very small, and which will not only affect the screening efficiency of the sample but also increase the dust concentration in the environment due to the too small cadmium powder particles after the screening. It is also a certain health hazard to the sample preparation personnel. Another important reason is that the pyrotechnic powder with a particle size of less than 100 mesh has flammability and is easily ignited in the air.

4.2 Advantages

A method for rapid and semi-quantitative determination of cadmium content in pyrotechnic samples of fireworks and firecrackers based on energy dispersive X-ray fluorescence spectrometry, the advantages of which are as follows: (1) The method is simple to operate, and the method can be repeatedly called for testing. Only one new test method needs to be built before the sample test, and after the method is established, the test can be repeated at different times without re-establishing the test method for each test. After the establishment of the new test method, the entire test process only includes three steps: sample preparation, sample loading into the sample cup and on-board testing. (2) The analysis time of the method is extremely short, and after the preparation of the sample, the entire measurement process only takes about 1 minute. (3) The method has low labor intensity and is not demanding to the operator. (4) The accuracy is good, the precision is high, and the false positive rate is low.

4.3 Method validation test

Because the standard of pyrotechnics with a certain amount of cadmium content can not be found in the market, and the physical form of black powder is similar to that of pyrotechnics, the reference material for the different cadmium content of black powder as the matrix configured with the standard material of cadmium can be tested as the samples. By comparing the correspondence between the cadmium content of different pyrotechnic reference materials and their corresponding characteristic fluorescence intensity values, the general correspondence between the cadmium content in the pyrotechnic composition and its corresponding characteristic fluorescence intensity would be inferred. The numerical relationship between the fluorescence intensity value and the content value of the cadmium element in the samples can be seen in Table 2.

Sample No.	1	2	3	4	5	6
Cd content(%)	0	10	30	50	80	99.9
Cd Fluorescence	0	4262	13121	18954	35473	39987
intensity						
values(cps/mA)						
Ratio	0	426.2	437.4	379.1	443.4	400.3

 Table 2. The numerical relationship between the fluorescence intensity value and the content value of the cadmium element

It can be seen from Table 2 that: Firstly, When the sample does not contain cadmium, the fluorescence intensity value of the characteristic line of the cadmium element in the corresponding method is also zero. Secondly, Observing the point where the cadmium content differs greatly, the fluorescence intensity value of the corresponding characteristic line is enhanced with the increase of cadmium content, which is positively correlated, but not strictly proportional. The main reason is that the matrix effects of each element in the sample on the cadmium element is more obvious due to the increase of the content of cadmium, and the direct effects of these matrix effects will increase the corresponding difference in the fluorescence intensity value of the characteristic line of the cadmium element. Thirdly, The mass percentage (%) of cadmium in the sample is positively correlated with the ratio of the corresponding cadmium element characteristic line fluorescence intensity value (cps/mA), and the ratio is within a range of 1:415±10% (specific value 1:379.1~443.4).

4.4 Method repeatability test

The purpose of the method repeatability test is to further confirm the corresponding characteristic fluorescence intensity values of cadmium elements when cadmium has a mass percentage of 1% in different types of pyrotechnic agents. The samples of pyrotechnic reference materials with cadmium content of 1% were determined by using different types of pyrotechnic agents as substrates. The specific test results are shown in Table 3.

 Table 3. Fluorescence intensity values of characteristic lines of cadmium in different pyrotechnic reference materials with cadmium content of 1%

Sample No.	1	2	3	4	5
Pyrotechnic effects	Red	Blue	Yellow	Green	White
Cd Fluorescence	405	431	388	424	365
intensity values(cps/mA)					
Ratio	405	431	388	424	365

It can be seen from the test data of Table 3 that when the content of cadmium carbonate in the sample of different pyrotechnic composition is 1%, the fluorescence intensity value of the corresponding cadmium element characteristic line ranges from 365 to 431 cps/mA. The ratio of the content of cadmium in the pyrotechnic composition to the corresponding fluorescence intensity value of the cadmium element is also positively correlated and fluctuates within a certain range (specific values are 1:365~1:431). Since it is a semi-quantitative test, this proportional coefficient can be approximated by an integer multiple of 1:360~430.

5. Conclusion

This article discloses a method for quickly and semi-quantitatively detecting cadmium content in pyrotechnic blind samples of fireworks and firecrackers based on energy dispersive X-ray fluorescence spectrometer (EDXRF), and the method has the advantages of simple operation, short detection period, good stability, good repeatability and high credibility.

References

[1] Leif, H. C.; Allan, A. Determination of sulfur and heavy metals in crude oil and petroleum products by energy-dispersive x-ray fluorescence spectrometry and fundamental parameter

approach.Anal. Chem., 53(12), pp. 1788-1792, 1981.

[2] Fei, H; Pierre, J. V. E. General approach for quantitative energy dispersive x-ray fluorescence analysis based on fundamental parameters. Anal. Chem.,63(20),pp.2237-2244,1991.

[3] Standard test method for determination of lead in paint layers and similar coatings or in substrates and homogenous materials by energy dispersive x-ray fluorescence spectrometry using multiple monochromatic excitation beams,ASTMF2853,American Society for Testing and Materials Publications, 2010.

[4] Standard test method for analysis of uranium and thorium in soils by energy dispersive x-ray fluorescence spectroscopy, ASTM C1255, American Society for Testing and Materials Publications, 2011.

[5] Duan Tiyu, Li Suqing, et al. Jewellery,Determination of precious metal content - Method using x-ray fluorescence spectrometry,China National Recommended Standard: GB/T 18043-2013, China Standard Press: Beijing, 2014.

[6] Leoni. L,Saitta. M, X-ray fluorescence analysis of powder pellets utilizing a small quantity of material, X-ray Spectrom, 3,pp.74-77,1974.

[7] Rose W.I.,Bornhorst T.J.,Sivonen S.J., Rapid high-quality major and trace element analysis of powdered rock by x-ray fluorescence spectrometry, X-ray Spectrom, 15,pp.55-60,1986.

[8] Gy,Pierre M,The analytical and economic importance of correctness in sampling, Anal.Chim.Acta, 190,pp.13-23,1986.

[9] Zhou Tonghui, Wang Erkang, Lu Wanzhen ect. Handbook in analytical chemistry(second edition), basic knowledge of and safety knowledge(the first volume), China Chemical industry press:pp.568-580,1997.